

EVALUATION OF YORK® INFRARED ANALYSER ON FAT AND PROTEIN CONTENT

YORK® mid-infrared spectrophotometer permits to determine fat content, protein matter and lactose, in milk and dairy products. CECALAIT physico-chemical laboratory conducted evaluation assays on apparatus stability, contamination between samples, linearity, repeatability and accuracy. This study concerned fat and protein criteria. The results obtained pertaining to stability, contamination, linearity, calibration, repeatability and accuracy, comply with the IDF standard 141C:2000 "Whole milk – determination of milk fat, protein and lactose content (Guidance for the operation of mid-infrared instruments)", and the CNIEL "Manuel d'utilisation des analyseurs infrarouges pour les laboratoires interprofessionnels et de contrôle laitier" requirements (= *User's guide to infrared analysers for interprofessional and milk control laboratories*).

The York® mid-infrared spectrophotometer (2-10 µm) is a manual apparatus (without conveyor) with an analytical rate of 120 samples/h. It is designed for the determination of the composition of certain constituents in milk and dairy products. It is manufactured by On-Line Instrumentation (GB) and marketed by Swantech.

The apparatus works using a filter mono-beam infrared system. It is also equipped with a sample re-heating system. From a MLR* calibration, it determines the common constituents (fat, protein and lactose). It is coupled with a micro-computer which deals with running and signal processing.

TESTS

CECALAIT physico-chemical laboratory conducted the evaluation tests. They were concerned with reference and infrared analyses for fat and protein. The tests took place from December 2002 to June 2003 and dealt with evaluating the apparatus stability. Then, after calibration of the device, the following characteristics were evaluated : contamination between samples, linearity, repeatability and accuracy.

While being evaluated, certain improvements were carried out on the apparatus. Among others, an on-line reheater was fitted. The evaluation criteria of estimated parameters were taken from IDF 141C:2000 "Whole milk – determination of milk fat, protein and lactose content (guidance for the operation of mid-infrared instruments) and from the CNIEL's Guide.

STABILITY

The evaluation of stability was performed by analysis of duplicate milk samples every 20 minutes for half a day

(representing 7 measurement cycles) according to the actual working conditions of a milk payment laboratory.

In order to evaluate the apparatus stability, repeatability and reproductibility parameters were calculated for each analytical criterion.

Criterion	M (g/l)	Sr (g/l)	SR (g/l)	Sr (%)	SR (%)	r (g/l)	R (g/l)
Fat	37.78	0.09	0.25	0.25	0.75	0.26	0.78
Protein	32.09	0.11	0.19	0.40	0.63	0.36	0.56

M : mean

Sr and *SR* : standard deviation of repeatability and reproductibility

Sr% and *SR%* : relative deviation of repeatability and reproductibility

r and *R* : maximal difference of repeatability and reproductibility within 95% of cases

Concerning the fat and protein criteria, the standard deviation of repeatability and reproductibility (SR) observed with the control session complies with those deduced from the IDF standard 141C:2000 ($SR \leq L / 2,58$ soit 0.27 ; L = control card limit at 99 % equals 0,7 g/l).

CARRY-OVER EFFECT

The carry-over effect was evaluated by analysing one batch of milk and distilled water, 20 times for fat and protein content, in the following sequence : MILK (L1) – MILK (L2) – WATER (W1) – WATER (W2). The volume pumped was 12 ml.

The carry-over effect was estimated with the following formula :

$$Tc(\%) = 100 \times [\sum W1 - \sum W2] / [\sum M1 - \sum M2]$$

Criterion	Mean level (g/l)	Carry-over (%)
Fat	37.88	0.00
Protein	31.83	0.00

The carry-over effect between successive samples is lower than 1% for each criterion tested (acceptability limit according to the IDF standard 141C:2000 and CNIEL Guide).

LINEARITY

In every case, volume/volume dilutions were performed by corrected weighing of the volumic mass. This corresponds to the infrared spectrophotometry dosage and to the French reference measurements which are equally volumic.

* Fat : a set of 14 evenly distributed milks set, from 0 to 120 g/l, was elaborated using a mixture of cream and skimmed milk.

* Protein : a set of 14 evenly distributed milks set, from 0 to 80 g/l, was elaborated using a mixture of proteinic retentate and filtrate obtained by tangential ultrafiltration (cut off level : 10 KD).

Each set was evaluated in duplicate, following increasing, then decreasing levels of fat or protein matter.

* Fat :

Figure 1 : Linearity deviation distribution (whole range)

(see at the end of this article)

Figure 2 : Linearity deviation distribution (adjusted range from 22 to 56 g/l)

(see at the end of this article)

The above graphs represent the deviation distribution of linearity according to the different levels obtained from dilutions of cream in skimmed milk over the whole range (figure 1) or after digital adjustment (figure 2) on the apparatus calibration range (from sampl 3 to sample 7 or from 22 to 56 g/l)

The residual standard deviations of regression are :

Type of regression	Whole range (g/l)	Calibration range (g/l)
1 st order	0.51	0.15
2 nd order	0.23	/
3 rd order	0.15	/

It can be noted that the apparatus is not linear over the whole range tested : after adjustment between 22 and 56 g/l, the residual deviation is about -0.8 g/l to 70 g/l and from -2.4 g/l to 100 g/l. However, it is possible to observe two areas of linearity (between 20 and 60 g/l and between 60 to 100 g/l).

It is also possible to improve the linearity of the apparatus by applying a correction using a 3rd order polynomial.

* Protein content :

Figure 3 : Linearity deviation distribution

(see at the end of this article)

Figure 4 : Linearity deviation distribution (adjusted range from 25 to 40 g/l)

(see at the end of this article)

The above graphs represent the deviation distribution of linearity according to the different levels obtained with ultrafiltration retentate and filtrate dilutions over the whole range (figure 3) or after digital adjustment (figure 4) on the

apparatus calibration range (from sample 4 to sample 7 or from 25 to 40 g/l).

The residual standard deviations of regression are :

Regression type	Whole range (g/l)	Calibration range (g/l)
1 st order	0.32	0.03
2 nd order	0.07	/
3 rd order	0.07	/

The results showed a minor linearity defect of the instrument over the whole range tested. After adjustment between 25 and 40 g/l, the residual deviation is about 0.5 g/l to 50 g/l and 1.9 g/l to 80 g/l. Two areas of linearity can also be observed (between 0 and 40 g/l and between 40 to 80 g/l).

It is also possible to improve the linearity of the apparatus by applying a correction using a 2nd polynomial.

For fat and protein contents, the linearity of the instrument is satisfactory in the range corresponding to mean contents (FM : 20 to 55 g/l ; PM : 20 to 40 g/l). However, it is possible to widen the usable area, by using an equation to correct the results ; either a 3rd order (FM) or a 2nd order polynomial (PM).

CALIBRATION

Calibration was evaluated using a range of 13 recombined milk samples in an orthogonal matrix of fat and protein contents complying with the IDF standard 141C:2000 requirements. Each sample was analysed in successive duplicates from 1 to 13. A control was set at the beginning and at the end in order to ensure stability. Calibration was performed on 03/06/2003.

The table below recapitulates the results obtained using simple and multiple linear regression.

Criterion	Y=b.x + a			
	b	a	Sd (g/l)	Sy,x (g/l)
Fat				
Protein	1.002	-0.14	0.19	0.19
	1.005	-0.17	0.09	0.09

Criterion	Y=a.C1 + b.C2 + c.C3 + d				
	c	b	z	d	Sy,x (g/l)
Fat					
Protein	-0.01	0.00	1.00	0.57	0.21
	0.01	0.00	1.00	-0.45	0.10

b and a : slope and bias of the simple linear regression

Sd : standard deviation

Sy,x : standard residual deviation

c,b,a,d : multiple linear regression factors

(FM : REF = a.FM + b.PM + c.LA + d;

PM : REF = a.PM + b.FM + c.LA + d)

Results of the simple linear regression are satisfactory since the residual standard deviations are lower than the acceptability limits set by the CNIEL Guide. (FM 20 g/l ; PM 0.15 g/l). There are no significant interactions between components.

REPEATABILITY

The repeatability was evaluated using 87 herd milk samples from the Jura county. Samples contained bronopol (0.02 %).

Dosages were performed in consecutive duplicates for each sample on 04/06/2003.

Criterion	n	Min (g/l)	Max (g/l)	M (g/l)	Sx (g/l)	Sr (g/l)	Sr (%)	R (g/l)
Fat	87	33.18	40.64	36.92	1.89	0.07	0.19	0.20
Protein	87	30.16	37.16	32.74	1.17	0.06	0.19	0.17

n : number of results

min and max : minimum and maximum values

M : mean results

Sx : standard deviation of results

Sr et Sr (%) : relative and absolute standard deviation of repeatability

r : maximal deviation of repeatability in 95% of cases

For fat matter and protein matter, the analyser shows a repeatability that complies with the requirements of the IDF standard 141C:2000 and with the CNIEL Guide ($r \leq 0.4$ g/l or $Sr \leq 0.14$ g/l).

ACCURACY

Samples were identical to those used in the evaluation of repeatability. Infrared analyses were performed using successive duplicates (means of duplicates were used for statistical treatments), a control milk was set every 20 samples in order to check the signal stability. The evaluation concerned values obtained after calibration. Analyses were performed on 04/06/2003.

Reference analyses :

* **Fat content** : (Gerber method) ; single test performed but confirmed in the case of too much residue)

NF standard V 04-210:2000 – Lait – Détermination de la teneur en matière grasse – Méthode acido-butyrométrique.
= Milk – Determination of fat content – Gerber method –
(This standard is not equivalent to ISO 2446:1976 though it deals with the same topic).

* **Protein content** : (in duplicate)

IDF standard 98A:1985 – Milk – Determination of protein content – Amido black dye-binding method (routine method)

	FM	PM
n	86	87
Min (g/l)	33.00	29.87
Max (g/l)	40.00	37.12
Y (g/l)	36.69	32.29
X (g/l)	36.91	32.74
Sy (g/l)	1.74	1.26
d (g/l)	0.22	0.44
Sd (g/l)	0.40	0.22
Sy,x (g/l)	0.35	0.21
b	0.901	1.056
a	3.42	-2.26

n, min, max : number of results, minimum and maximum values

Y, X : mean results obtained by reference and instrumental method

Sy : standard deviation of results obtained by the reference method

d, Sd : mean and standard deviation of deviations

Sy,x : residual standard deviation

b, a : slope and bias of the simple linear regression

The table and graphs above show the results obtained :

(see at the end of this article)

* **for fat matter**, the mean result of deviations obtained is 0.22 g/l and the residual standard deviation of linear regression is equal to 0.35 g/l. The slope (0.901) and bias (3.42) are different from 1 and 0 ($P=1\%$), respectively.

* **for protein matter**, the mean result of deviations obtained is 0.44 g/l and the residual standard deviation of linear regression is equal to 0.21 g/l. The slope (1.056) and bias (-2.26) are respectively different from 1 and 0 ($P=1\%$).

For fat and protein matter, the slope and bias deviations are particularly high and could be explained by :

- the sample composition for calibration (they came from a different geographic area)
- the gap between the age of the calibration range (June 2003 performed on 07/05/2003) and herd milk analysis (04/06/2003)
- the low amplitude of FM and PM rate

Instability of the analyser is excluded, as the results obtained from the control milk used are similar (the control milk was identical during calibration and evaluation of accuracy).

However, regarding stability, carry-over effect, linearity, calibration, repeatability and accuracy, the analyser complies with the requirements of IDF 141C:2000 set at 0.7 g/l for herd milk, the residual standard deviation of regression values.

CONCLUSION

The results regarding stability, carry-over effect, linearity, calibration, repeatability and accuracy, comply with the requirements of IDF 141C:2000 and with the CNIEL "Manuel d'utilisation des analyseurs infrarouges pour les laboratoires interprofessionnels et de contrôle laitier" (= User's guide to infrared analysers for interprofessional and milk control laboratories)

X. QUERVEL, Ph. TROSSAT

ABBREVIATIONS

CNIEL : Centre National Interprofessionnel de l'Economie Laitière = Interprofessional Centre for the Dairy Economy
FM, PM : Fat matter, protein matter
LA : Lactose

BIBLIOGRAPHY

Standards :

- NF V 04-210:2000 – Milk – Determination of fat content – Gerber method (*this standard IS NOT equivalent to ISO 2446:1976 though it deals with the same topic*)
- IDF 98A:1985 – Milk – Determination of protein content - Amido black dye-binding method (routine method)
- IDF 141C:2000 – Whole milk – Determination of milk fat, protein and lactose content (Guidance for the operation of mid-infrared instruments)

Guide and reports :

- CNIEL and Institut de l'Elevage, Manuel d'utilisation des analyseurs infrarouges pour les laboratoires interprofessionnels et de contrôle laitier, (référence : CNIEL PROC IR-04-05/00), 2000, 48p.
- QUERVEL X, TROSSAT P. Rapport d'évaluation de l'analyseur infrarouge YORK®, CECALAIT, 18/08/2003, 8 p
- .