

EVALUATION OF THE LACTOSCOPE® INFRARED ANALYSER

Developed by Delta Instruments, the Lactoscope® is a mid-infrared Fourier transform analyser for the determination of fat, protein and lactose content in milk, as well as other criteria such as urea content equivalent (FNCV) and determination of freezing point (FP) equivalent.

The phase 1 assay took place from July to September 2001. CECALAIT evaluated the analytical and instrumental characteristics of the Lactoscope® for the determination of fat and protein content, and FP. Its basic characteristics, stability and carry-over effect, appear to be very satisfactory. Over a range of "usual" levels, it is also satisfactory for the adjustment of linearity proposed by the developer. Finally, the repeatability and accuracy values, for individual or herd milk, appear to comply with the regulations and standardization requirements.

PRINCIPLE AND DESCRIPTION

The Lactoscope® uses a mono-beam infrared system based on the Fourier transform technic and it includes an interferometer. It is coupled with a micro-computer which deals with running and signal processing, calibration and samples.

The calibration of the apparatus is carried out using a multiple linear regression calculation, from absorbances obtained at various wavelengths (usually 3 wavelengths for fat and 4 for protein, according to the IDF standard 141). This calibration is suitable for the prediction of the "usual" parameters : fat, protein and lactose, but also for the prediction of the "urea content" equivalent or for the "freezing point" equivalent.

TESTS PERFORMED

CECALAIT physico-chemistry laboratory conducted all the evaluation tests, which concerned the reference and infrared analyses for fat and protein content and FP.

At a rate of 360 determinations / h, the tests dealt with the evaluation of stability, carry-over effect, linearity, repeatability and accuracy of the apparatus.

The evaluation criteria of estimated parameters were taken from the IDF standard 141 C : 2000 - Whole milk - Determination of milkfat, protein and lactose content (guidance for the operation of mid-infrared instruments) or from the "CNIEL - Manuel d'utilisation des appareils infrarouge dans le cadre du paiement du lait en France " (= User's guide to infrared analysers for milk payment in France).

STABILITY

The evaluation of stability was performed by analysis of three milk samples covering a range of normal measurements for fat and protein content. The milk samples were analysed automatically, in duplicate, every 15 minutes over half a day (representing 14 measurement cycles)

according to the actual working conditions of a milk payment laboratory.

In order to evaluate the apparatus' repeatability and reproducibility, parameters were calculated for each analytical criterion. The more important results are shown below.

Table 1 - Summary of the reproducibility values obtained when evaluating the stability of the Lactoscope®

Levels	Fat SR (g/l)	Protein SR (g/l)	FPD R-(m°C)
1	0.08	0.06	3.87
2	0.09	0.05	3.04
3	0.10	0.06	2.95

SR : absolute standard deviation of reproducibility

R : maximum reproducibility difference in 95 % of occurrences

Results

Concerning the fat and protein criteria, the mean daily values of the standard deviation of reproducibility observed, comply with those deduced from the "Manuel d'utilisation des appareils infrarouge dans le cadre du paiement du lait en France " (SR < L / 2,58 ; L : control card limit at 99 % equals 0,7 g/l).

Concerning the FP equivalent, in the absence of standardized values, the IDF 108 B reproducibility value was used (R= 5 m°C). The reproducibility values obtained here are all lower than the IDF value.

CARRY-OVER EFFECT

The carry-over effect was evaluated by analysing automatically one batch of milk and distilled water, 20 times in the following sequence : MILK (M1) - MILK (M2) -MILK (M3).

The measurements concerned fat and protein content, at 3 levels for each criterion, respectively (20, 20) for milk 1; (40,30) for milk 2; (60, 40) for milk 3.

NB : In order to permit the start the analysis, NaCl was added to the water sample to give a final concentration of 0,5 %.

The carry-over effect (Tc) was estimated with the following formula :

$$Tc (\%) = [\Sigma (\text{water1}) - \Sigma (\text{water2})] / (\Sigma (\text{milk1}) - \Sigma (\text{milk2}))] \times 100$$

Results

Contamination between successive samples appears to be from 0,53 % to 0,70 %, whatever the constituent and the level tested.

These levels comply with the 1 % acceptability limit, applied to methods used for the rapid determination of milk quality, in the context of milk payment and milk control.

LINEARITY

For fat content, the linearity was evaluated using a set of 14 evenly distributed milk samples, from 0 to 128 g/l, elaborated using a mixture of cream and skimmed milk.

For protein content, the linearity was evaluated using a set of 14 evenly distributed milk samples, from 0 to 83 g/l elaborated using a mixture of proteinic retentate and filtrate obtained by tangential ultrafiltration (cut off level : 10 KD).

The results showed that the adjustment proposed by the manufacturer is satisfactory :

- in the range corresponding to the protein criterion
- in the range corresponding to the calibration of the apparatus, (22 to 56 g/l).

However, it is possible to widen the usable area, by using a 3rd order polynomial equation to correct the results.

REPEATABILITY

For fat and protein content, the repeatability was evaluated using 131 individual milk samples from 7 breeding farms in the Jura county and 58 herd milks.

Milk samples covered a range of values which were from 17 to 85 g/l for fat content and from 25 to 43 g/l for protein content.

For the FP equivalent, the repeatability was evaluated using 58 herd milk samples.

Samples contained bronopol (0.02 %). Measurements were performed in consecutive duplicates analysing

automatically each set of samples in the following sequence :

Set 1 rep 1 - Set 1 rep 2 - Set 2 rep 1 - Set 2 rep 2 - ...
Set n rep 1 - Set n rep 2.

Tables 2 and 3 show the results obtained.

Sx : standard deviation of results
Sr et Sr (%) : relative and absolute standard deviation of repeatability
r : maximal deviation of repeatability in 95% of cases
n : number of results
min and max : minimum and maximum values
M : mean results

Table 2 : Evaluation of repeatability for individual milk

INDIVIDUAL MILK					
Criterion	n	M	Sr	Sr (%)	r
Fat	131	41.37	0.08	0.20	0.23
Protein	131	33.79	0.07	0.22	0.20

Table 3 - Evaluation of repeatability for herd milk

HERD MILK						
Criterion (g/l)	n	M	Sx	Sr	Sr (%)	r
Fat	58	37.89	1.94.	0.09	0.24	0.25
Protein	58	32.95	1.26	0.07	0.20	0.19
FPD (m)C x - 1	58	504	10.0	1.6	0.31	4.4

Results

Tables 2 and 3 show that for fat and protein matter, the analyser offers a repeatability that complies with the requirements of the IDF standard 141 C : 2000 (Sr = 0.14 g/l and r =0.4 g/l).

Concerning the FP equivalent, the repeatability value obtained is close to the repeatability of the Thermistor cryscope reference method (Sr = 1.4 m°C).

ACCURACY

The accuracy was evaluated using two types of sample:

- 100 individual milk samples (out of 131) from 7 breeding farms in the Jura county, to which bronopol was added (0.02 %), for the evaluation of fat and protein content,
- 58 herd milk samples from the Franche-Comté region, to which bronopol was added (0.02 %), for the evaluation of fat, protein and FP equivalent.

The reference methods used are those used in the context of milk payment :

- Fat - GERBER method (NF V 04 210 AFNOR standard), single analysis but with confirmation when residues are too important.
- Protein - Amido Black method (NF V 04 216 AFNOR standard), with duplicate analyses.
- FP equivalent : plateau seeking method with use of a thermistor (NF V 04 205 AFNOR standard).

Results

Tables 4 and 5 show the results obtained, respectively for fat and protein content, using individual milk and herd milk samples. Table 6 shows results corresponding to FP equivalent measurement, using herd milk samples only. (see below)

Table 4 - Accuracy of the Lactoscope® for fat

	INDIVIDUAL MILKS	HERD MILKS
n	100	58
Y (g/l)	38.82	37.71
X	38.85	37.89
Sy	8.40	1.93
d	0.03	0.18
Sd	0.46	0.34
Sy,x	0.433	0.344
b	0.981	0.980
a	0.72	0.56

Table 5 - Accuracy of the Lactoscope® for protein

	INDIVIDUAL MILKS	HERD MILKS
n	100	58
Y (g/l)	33.16	33.11
X	32.72	32.95
Sy	2.61	1.28
d	- 0.44	- 0.16
Sd	0.38	0.21
Sy,x	0.382	0.207
b	0.991	1.007
a	0.75	- 0.08

Table 6 - Accuracy of Lactoscope® for FPD equivalent

	HERD MILKS
n	55
Y (g/l)	516
X	504
Sy	10
d	- 13
Sd	3
Sy,x	2.8
b	0.927
a	49

Keys for tables 4, 5 and 6

n : number of samples

Y : mean of the results using the reference method

X : mean of the results using the Lactoscope®

Sy : standard deviation of the results using the reference method

d : mean of the differences Lactoscope® - reference method

Sd : standard deviation of the difference

Sy,x : residual standard deviation

b : slope of the linear regression equation

a : **point 0 ordinate = intercept?**

CONCLUSION

Stability and carry-over effect appear very satisfactory. For the range of "usual" levels, it is also satisfactory for the adjustment of linearity proposed by the manufacturer. Finally, the repeatability and accuracy values, for individual or herd milk, appear to comply with the regulations and standardization requirements.

Authors : X. QUERVEL, P. TROSSAT

Translated by I. BECAR

Emended by Helen LAMPRELL

Bibliography

AFNOR standards : V 04-205 - janvier 1990 ; V 04-210, septembre 2000 ; V 04-216, décembre 1985.

CNIEL, Manuel d'utilisation des appareils infrarouge dans le cadre du paiement du lait en France " (= User's guide to infrared analysers for milk payment in France), Réf : Proc IR-04-05/00.

IDF 108 B : 1991 - Milk - Determination of freezing point
(Thermistor cryoscope method)

IDF 128 A : 1999 - Milk - Definition and evaluation of the
overall accuracy of indirect methods of milk analysis,
application to calibration procedure and quality control in
the laboratory

IDF 141 C : 2000 - Whole milk - Determination of milkfat,
protein and lactose content (guidance for the operation of
mid-infrared instruments)

QUERVEL X, TROSSAT P, Rapport préliminaire
d'évaluation du Lactoscope®, CECALAIT, 2001, 13 p.

Abbreviations

AFNOR = French body for standardization

CNIEL : Centre National Interprofessionnel de l'Economie
Laitière = Interprofessional Centre for the Dairy
Economy

IDF : International Dairy Federation

FNCV : Feed Nitrogen Conversion Value

FPD : Freezing Point Determination

